

CERTIFICATE OF CONFORMITY

Product: Analysis equipment (fluid sample handler)

Tested by request of Gold Standard Diagnostics
2851 Spafford Street
California 95618 DAVIS
United States of America

Manufactured at (name and place) Gold Standard Engineering BVBA
Esperantolaan, 7A
3300 TIENEN
Belgium

Trade mark (if any): GOLD STANDARD DIAGNOSTICS

Model/Type Ref: ThunderBolt EIA

Ratings and principal characteristics: 100-240 V, 50-60 Hz, Max 120 W, class I, 30°C max.

Additional information (if any) :

A sample of the product has been tested and found to be in conformity with :

EN 61010-1:2001
IEC 61010-1:2001-02
US and CA national deviations

as shown in the test report 586125.01

This Certificate of Conformity is the result of testing a sample of the product submitted, in accordance with the provisions of the relevant specific standard. It does not imply an assessment of the whole production. Conformity of the produced products with the specimen tested remains on the full responsibility of the manufacturer.

Brussels, 26/01/2011

C. LANA
Certification Manager.